

SARAVANA MEDICO

Shop No. 41/50, Good Shed Road, 4th Street, Adambakkam, Chennai - 600 088
E-mail : saravanamedico2019@gmail.com, **Mobile : 7358077155**

Certificate No.: E131221-1-1

Name & Address of the Customer

**M/s. Upgraded Primary Health Center,
Periyar Nagar Avadi**

SRF No. : E131221-1

Received on : 25.09.2022

Date of Calibration : 30.09.2022

Recom. Due Date : 29.11.2022

Details of Test Instrument

Description	: Micropipette	Range	: 1 ml
Make	: ---	Operating Range	: ---
Model No. / Type	: ---	Resolution	: ---
Serial No.	: ---	Accuracy	: As per specification
ID No. / Asset No.	: --	Location	: ---
Calibrated at	: Lab	Condition On Receipt	: Satisfactory
Work Instruction No.	: EMTPL/WI/MECH-14	Based on Standard	: ISO 8655-6 : 2002
Environmental Condition	: Temperature: 23±1.5 °C	Relative Humidity	: 50±10 %RH

Details of Master Instrument

Description	ID No.	Traceability	Valid till
Electronic Semi Micro Balance	EMTPL-EWB-01	2022-23/EQN/210252-01	29.11.2022

The Masters are traceable to National/International Standard

CALIBRATION DATA MECHANICAL CALIBRATION (Volume)

Test Reading	Standard Reading	Deviation	Expanded Uncertainty (±)
ml	ml	ml	ml
1	0.996	-0.004	---

Remarks :

1. This certificate refers to the values for the above instrument at the time of calibration and under the above stated conditions.
2. The Calibration Certificate shall not be reproduced except in full without the written approval of EMTPL
3. The Reported Expanded Uncertainty is at 95.45% Confidence level., coverage factor k=2



Calibrated by :
Quality Manager



Authorised by :
Technical Manager