

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023

| | | | |
|-------------------|-------------------------------|----------------------------|------------|
| Lab Name | Govt. UPHC, Anuppanadi | Lab No | 10794 |
| Constituent Group | Chemistry III | Date of Result Entered : | 20/09/2023 |
| PT item | Lyophilized human serum based | Date of Report Published : | 05/10/2023 |

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1870 | 310.86 | 4.82 | 14.98 | 321 mg/dL | 0.68 | 0.69 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.29 | 8.14 | 0.10 | 1.12 mg/dL | -1.62 | 0.00 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1783 | 113.62 | 3.80 | 4.32 | 107 mg/dL | -1.53 | 0.20 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. UPHC, Anuppanadi
 THE MEDICAL OFFICER, AYYANAR KOVIL STREET, ANUPPANADI
 Madurai-Corp
 TAMILNADU625009

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****