

CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical officer Government Primary Health Centre Mimi-Pudhukottai(dist)	Certificate no : Rmtl/01/319107848/46 Date of calibration :05/08/2023 Suggested calibration due : 04/08/2024 Customer po .reference.& date:
---	--

DESCRIPTION OF DUC

Name Of The Duc: pipette Make : THERMOSCIENTIFIC Model : Finnpiquette Sl.No : RW10692 Range/Size : 10-100 micro liter	Receiving Condition Of Duc :Good Ref. Document : As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
--	--

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

DETAILS OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-202
Micro balance	mettle toledo/XP26PC	QP-01	B115129039/CC-2523	1-6-202
Universal Measuring Instrument With Sensor For Environmental Parameter	ALMEMO/2590-4AS	QP-10	H19070544/CC-2052	1-6-202

CALIBRATION FIELD: WEIGHT Calibration

Sl.no	Range	Master readings micro litter	Calibrated values micro litter	Error observed micro litter	Uncertainty
1	10-100 micro litter	10	9.9	0.1	0.055
2	10-100 micro litter	50	49.9	0.1	0.055
3	10-100 micro litter	80	79.9	0.1	0.055
4	10-100 micro litter	90	89.9	0.1	0.055
5	10-100 micro litter	100	99.9	0.1	0.055


remarks

uncertainty estimated for a confidence level of 95% at the k factor=2


the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration

DUC –device under calibration

Calibration label /status : Calibration result pass & fail are indicated


 CALIBRATION ENGINEER




 AUTHORIZED SIGN

This Report Refers Only To The Sample Submitted And
 May Not Be Reproduces In Full Or In Part, Without
 Written Permission From Roots Metrology & Testing
 Laboratory, Coimbatore, cell:9585840192