



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. Block PHC, Konganapuram  
 Constituent Group: Chemistry III  
 PT Item: Lymphitized human serum based

Lab No: 7467  
 Date of Result Entered: 15/08/2023  
 Date of Report Published: 24/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Target Value	SD	CV
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.28	190 mg/dL	1.96	9.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.60 mg/dL	0.07	9.09
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.54	125 mg/dL	2.40	9.20

SD Range	Interpretation
Below 1.00 to <1.00	Excellent.
Below 1.01 to <2.00	Good.
Below 2.01 to <3.00	Accept with caution. Warning Signal.
Below >3.00	Unacceptable performance. Action Signal.

LAB ADDRESS:  
 Govt. Block PHC, Konganapuram  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT,  
 Konganapuram  
 158844/2023/103

*Pradeep Chidambaram*