



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MARCH 2023



PC-102A

Lab Name: Govt. PHC, Nedimozhiyanur
 Constituent Group: Chemistry III

PT Item: Lyophilized human serum based

Lab No: 13077
 Date of Result Entered: 20/03/2023
 Date of Report Published: 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	88.71 mg/dL	-0.41	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.33 mg/dL	0.00	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	99.31 mg/dL	-2.00	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Nedimozhiyanur
 Rattanai road
 Nedimozhiyanur
 TAMILNADU604306

MEDICAL OFFICER
Primary Health Centre,
Nedimozhiyanur - 604 306

Pavala Christudas

Dr. Pavala Christudas
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Coordinator Contact Details:
 Email: ctingc@cmcvellore.ac.in
 Contact Number: 9416-2283102