

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address : M/s Government Primary Health Centre Pethappampatty-6422205	Report No	VCPL/2023/002-2
	Date of Receipt	23-Mar-2023
	Date of Issue	24-Mar-2023
	Calibration Date	23-Mar-2023
	Next Due Date	22-Mar-2024
Customer Reference : Work Order	Calibration at	On Site

Details of Device Under Calibration:

Description:	BIO CHEMISTRY ANALYSER	Make:	ROBONIK
Range:	-	Model/Type:	-
Resolution:	-	Condition of DUC:	Good
Serial Number:	ATCD3160321RBK	Location:	Laboratory
ID number:	-	Cal. At:	ONSITE

Environmental Conditions		Calibration Method Used	
Temperature	25.9°C	National/International standard:	IEC60601-1
Humidity	52 %Rh	Cal procedure no:	-

Master Details


SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyser	05H-0600	ACCS221497	Rigel Medical / 288+	9-Aug-23

Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES μA	ALLOWABLE LIMITS Ω
2	Earth Leakage	166	<5000 μA For B,BF,CF
		MEASURED VALUES μA	ALLOWABLE LIMITS μA
3	Enclosure Leakage	223	<500 μA For B,BF,CF

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

(T.SUDHAHAR)
CALIBRATION ENGINEER

Authorised by:


S. LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

