

HIN/MED/2023-2024/1109882
12th Aug 2023

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer ABX Pentra ES60 bearing serial number: 603PES14009 installed at NARAYANA Diagnostic Center Kanpur as calibrated on 12th Aug 2023

Calibrator : ABX MINOCAL
Lot No. : CX487
Expiry Date : 05th Sep 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 11th Aug 2024



Shrish Dixit
(Head- Products & Marketing)
For Horiba India Pvt. Ltd.