



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Narayana Diagnostics. 7/225A, Lower Ground Floor, Gutaiya, Swaroop Nagar, Beside Murari Lal Chest Hospital, Kanpur-208002.	Service request No. & date	A/02/23/08/23&23/08/2023
		ULR no.	CC223923000006965F
		Cert. No.	AT23000006365
		Date of Receipt of DUC	23/08/2023
		Date of calibration	23/08/2023
		Date of issue	24/08/2023
		Suggested due date	22/08/2024

Instrument Details			
Instrument name	Dig. Centrifuge	Location	-----
Make / Model No.	Remi / NEYA-2	Accuracy	-----
Range / Size	0 to 4000 RPM	Visual Inspection	OK
Least Count	10 rpm	Calibration Performed at	Site
I.D. No. / Sr. No.	ND/CNB/12 / ZJBN-05691		

Detail of reference standards & Major equipment's used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/A.C92792		
Certificate No.	QCTS/000246/10		
Calibration Validity	20/10/2023		
Calibration by	QCTS		

Environmental Condition	Temperature	20±2 °C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Calibration Results			
Serial No.	Standard Value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	501.5	502.0	±0.2 %rdg.
02.	1005	1000	
03.	2010	2020	
04.	2990	3000	
05.	3980	4000	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
Calibration Engg./TM)
(MAHPAL)
Form No.- QF-47



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End of report---