

## **ASIAN TECHNOLOGY**

(A HOUSE OF CALIBRATION)



## CALIBRATION CERTIFICATE

Customer name	M/S: Narayana Diagnostics.	Service request No. & date	A/02/23/08/23&23/08/2023
And	7/225A, Lower Ground Floor,	ULR no.	CC223923000006972F
Address	Gutaiya, Swaroop Nagar, Beside	Cert. No.	AT23000006372
	Murari Lal Chest Hospital,	Date of Receipt of DUC	23/08/2023
**	Kanpur-208002.	Date of calibration	23/08/2023
	pur =====	Date of issue	24/08/2023
		Suggested due date	22/08/2024

	Instrument Details		
Instrument name	Deep Freezer -20		
Make / Model No.	Blue Star / CHFSD100DHSW	Location	
Range / Size	-20 to 0 °C	Accuracy	
Least Count	0.1 °C	Visual Inspection	OK
I.D. No. / Sr. no.	ND/CNB/18 / G57877027369001016	Calibration Performed at	Site

Detail of reference sta	andards & Major equipment's used	
Equipment Name	SPRT With Indicator	
Make	Tempsans& Fluke	
Model / SR No.	PT-100X1 & 1586A / 05640	
Certificate No.	CC288822000001839F	
Calibration Validity	27/10/2023	
Calibration by	Sigma Test & Research Centre	

Environmental	Temperature	25±4°C	Calibration Reference	ITS:90
Condition	Relative Humidity	50±20 %	Calibration Procedure	CP-03T

**Calibration Results** 

Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor k=2)	
01.	-20.033	-19.9		
02.	-10.055	-9.9		
03.	-5.067	5.1	±0.3 °C	
04.	0.012	0.2		

## Remarks:

- (1) Standard equipment use for calibration are traceable to national/international standards.
- (2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factork = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (DB SAXENA) Form No.-QF-47



Approved By (DM/TM) (NEERA) TYAGI) Page No. 1 of 1