



# G MULTI SERVICES

75, Yashob Street, Gobichattipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.  
gmultiservicegobi@gmail.com. Call : 96897 20654.

Bio Medical Instruments Services

Er. A. Bathul Rasig DECE., BE (ICE)  
Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE  
MAKE : FINNPIPETTE F3  
MODEL : 10-100 $\mu$ l

PROTECT CLASS /TYPE : Risk Class II  
S. No. : AR/PI/001  
Location : GPHC, ARACHALUR

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	-	-
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test (a) Mean 99.6 $\mu$ l (b) A -0.6 $\mu$ l (c) Acc % -0.71 (d) CV% -0.36	OK OK OK OK	
4	Calibration Test min volume 10 $\mu$ l max Volume 100 $\mu$ l (ELECTRONIC BALANCEING TEST) (a) A+or - 1.5, 5 $\mu$ l (b) S 1.5 $\mu$ l max Volume 5 $\mu$ l (c) Acc+ or - 1.50, 0.50 (d) CV% 0.40, 0.30	Ok Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory		Issued By:	
Endorsed By: Research team head		User Department:	
Reviewed On: 02.10.23 Next Review : 01.10.24			



*(Handwritten signature)*  
 மருத்துவ அளவீடுகள்  
 அளவு அளவீடு கருவிகளின் தரப்படுத்தல்  
 அறிவுரை - 888 1111  
 எழுந்தருளியுள்ள இடம்