



G MULTI SERVICES

7B, Yashob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 482.
gmultiservicesgobi@gmail.com. Cell : 96597 20684.
Bio Medical Instruments Services

Dr. A. Iyathurai Ravi DECE., BE (ICE)
Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE
MAKE : NA
MODEL : 10-100 μ L

PROTECT CLASS /TYPE : Risk Class II
S. No. : AR/PI/003
Location : GPHC, ARACHALUR

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	-	-
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test (i) Mean 99.5 μ l (j) A -0.6 μ l (k) Acc % -0.71 (l) CV% -0.36	OK OK OK OK	
4	Calibration Test min volume 10 μ l max Volume 100 μ l (ELECTRONIC BALANCEING TEST) (i) A \pm or - 1.5, 5 μ l (j) \pm 1.5 μ l max Volume 5 μ l (k) Acc \pm or - 1.50, 0.50 (l) CV% 0.50, 0.30	Ok Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory		Issued By:	User Department:
Endorsed By : Research team head			
Reviewed On: 02.10.23 Next Review : 01.10.24			

(Handwritten signature)
மருத்துவ ஆய்வகம்
சுகாதார அமைச்சு, சென்னை
சென்னை - 600 034
பொது மருத்துவமனை