



Vanavil Calibrations

Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page : 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, THIRUVATTAR-629 183.	Certificate No:	VCPL/CC/28-24/16-001
	Date of Calibration:	28-09-2023
	Date of issue:	29-09-2023
	Date of Receipt:	28-09-2023
	Next cal Date:	27-09-2024

Details of Device Under Calibration			
Description	CENTRIFUGE	Make:	REMI
Range	3500 RPM	Model/Type:	-
Resolution	10 RPM	Condition of DUC:	GOOD
Serial No / Id.No	GPHC/LAB/CEN-01	Instrument Location:	LABORATORY

Enviromental Conditions		Method Used For Calibration	
Temperature	25.4 °C	Reference Standard No	SANAS TR4501
Humidity	53 % Rh	Calibration Procedure No	-


Details of Kereence Standard Equipment For Calibration				
S.No	Description	ID No / SL No	Certificate No	Validity
1	DIGITAL TACHOMETER	DTM-01	HT/CC/221202-03/005 SYSTEMS/ HTM 550	4-Dec-23

Calibration Result				
S.No	DUC Reading (RPM)	Standard Reading (RPM)	Error (RPM)	Measurement Uncertainty (RPM)
1	500	500.2	-0.2	12
2	1000	999.6	0.4	12
3	2000	1997	3	12
4	3000	3001	-1	12
5	3500	3497	3	12

Remarks:


- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


C.AKASH
(CALIBRATION ENGINEER)



Authorised by:


S.LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

vanavilcalibrations@gmail.com www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com