



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Arangadurgam  
 Lab No: 10646  
 Month: June  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 21/06/2023

Date of Report Published : 04/07/2023

| SL.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |       | Your Value   | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|--------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD    |              |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1823               | 286.63 | 3.17         | 14.82 | 282.2 mg/dL  | -0.30 | 0.69 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1848               | 2.57   | 6.64         | 0.17  | 2.51 mg/dL   | -0.35 | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1804               | 113.73 | 4.46         | 5.07  | 111.51 mg/dL | -0.44 | 0.24 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

| Links                          |                                 |                                    |  |
|--------------------------------|---------------------------------|------------------------------------|--|
| <a href="#">Modify Address</a> | <a href="#">Modify Methods</a>  | <a href="#">Submit Lab Results</a> | <a href="#">View Lab Monthly Summary</a> |
| <a href="#">Yearly Summary</a> | <a href="#">Change Password</a> | <a href="#">Home</a>               | <a href="#">Sign-out</a>                 |



Site Maintained by

**Swiftware Solutions**

Best Viewed at pixels 1024 x 768