

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12517 MEDICAL DEVICES Page No:1 of 1

Issue Date	29-09-2023
SRF No & Date	SRF/23/00538-0003 & 28-09-2023
Receipt Date	28-09-2023
Calibration Date	28-09-2023
Calibration Due	27-09-2024

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE,
 ARANGALDURGAM-635811.

Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : NA & PRITEST TOUCH
Range : MULTI	Sr No : ATCD3361220RBK
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature 25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 98	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B, BF, CF	Measured values in µA 195	Uncertainty in % (±) 7.4
3	Enclosure Leakage <500µA for B, BF, CF	Measured values in µA 232	Uncertainty in % (±) 7.2

REMARKS

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory
- 2 The user should determine the suitability of the instrument for its intended use
- 3 The recalibration interval should be determined on the user requirement.
- 4 The results stated in this certificate relate only to the item calibrated
- 5 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- 6 Equipment used for Calibration were calibrated & traceable to National & International Standards

Calibrated By

Authorised Signatory


 (Calibration Engineer)
 M DINESH


 Technical Manager
 C SHANMUGARAJ

Chief Executive

