

NEW LAB MONTHLY SUMMARY

Govt. PHC, Anumantharayankottai  
 12149  
 April  
 2023  
 Chemistry III

Click on the analyte to view Graphical Data  
 All Analyser Result   
 Date of Result Entered : 19/04/2023  
 Date of Report Published : 06/05/2023

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Lab Name  
 Lab No  
 Month  
 Year  
 Constituent  
 Group

Method / Principle Name  
 Analyzer Name  
 No of Participants  
 DV  
 Participants CV SD U  
 Your Value

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV SD U	Your Value
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic)	1832	119.45	6.04 7.21	120 mg/dL 0.08 0.34
2	CREATININE III	Jaffes Kinetic - Alkalline Picrate III	Robonik ( Semi automatic / Fully Automatic)	1820	2.16	8.09 0.18	2.3 mg/dL 0.80 0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic)	1750	111.61	5.19	107 mg/dL -0.80 0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

**MEDICAL OFFICER**  
 GOVT. PRIMARY HEALTH CENTRE  
 HANUMANTHAYARANKOTTAI - 624 002.  
 DINDIGUL.