

CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 28.09.2023
		Calibration Valid: 27.03.2024
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, THO.JEDARPALAYAM,RASIPURAM, NAMAKKAL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	REMI
ii	Model	C-852
iii	Serial No	CECL3783
iv	Location	LABORATORY

CALIBRATION WORKSHEET

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	1999	-1

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB



99400 28355, 9751206106

88833 15662



No.42/1, NavalIndia Road, Peelamedu,

CBE - 04, hbmedicalsyste@gmail.com