

1:53:56 PM

**NEW LAB MONTHLY SUMMARY**

Govt. PHC, Anumantharayankottai  
 12149  
 May  
 2023  
 Chemistry III

Chemistry III  
 group

[All Analyser Result](#) [Print](#)

Click on the analyte to view Graphical Data


Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	190.3 mg/dL	1.00	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.67 mg/dL	-0.68	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	120.5 mg/dL	1.01	0.27

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.
Beyond ±3.0		Unacceptable performance. Action Signal.

<b>Links</b>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Modify Address</a>	<a href="#">Submit Lab Results</a>
<a href="#">Yearly Summary</a>	<a href="#">Home</a>
<a href="#">Modify Methods</a>	<a href="#">Sign-out</a>
<a href="#">Change Password</a>	

  
**MEDICAL OFFICER**  
 GOVT. PRIMARY HEALTH CENTRE  
 HANUMANTHARAYANKOTTAI - 624 002,  
 DINDIGUL.