

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Anumantharayanankottal
 Lab No 12149
 Month June
 Year 2023

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Constituent Chemistry III
 Group

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/06/2023


Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	282 mg/dL	-0.31	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.43 mg/dL	-0.82	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	117 mg/dL	0.64	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 HANUNTHAYARANKOTTAL - 621002.
 DINDIGUL.