

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Anumanthararakkottai	Details About Robust Analysis
Lab No	12149	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Chemistry III		
Group		

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	183.5 mg/dL	0.88	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	118.1 mg/dL	0.93	0.22

SDI Range

Interpretation