



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNALQUALITYASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. UPHC, Bangalore Road
 Konavattam
 Constituent Chemistry III
 Group
 PT item Lyophilized human serum based

Lab No 9300
 Date of Result Entered : 19/08/2023
 Date of Report Published : 01/09/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	167.6 mg/dl	-0.89	0.42
2	CREATININE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.74 mg/dl	-0.17	0.01
3	CHOLETEROL III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dl	-0.39	0.22

SDI Range	Interpretation
Within ± 1.00 to ± 1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performan"ce. Action Signal.

LAB ADDRESS :

Govt. UPHC, Bangalore Road, Konavattam
 BANGALORE ROAD, KONA VATTAM
 Bangalore Road, Konavattam
 TAMILNADU 632013

Pamela Christudoss

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMCEQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed,
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 *****End of Report*****