



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNALQUALITYASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. UPHC, Bangalore Road,  
Konavattam  
Constituent Group Chemistry III

Lab No

9300

Date of Result Entered :

19/07/2023

PT item Lyophilized human serum based

Date of Report Published :

07/08/2023

SLN	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	110 mg/dl	0.11	0.28
2	CREATININE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.5mg	-0.94	0.02
3	CHOLETEROL III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dl	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

### LAB ADDRESS :

Govt. UPHC, Bangalore Road, Konavattam  
BANGLORE ROAD, KONA VATTAM  
Bangalore Road, Konavattam  
TAMILNADU 632007

*Pamela Christudoss*

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
contact Number: 0416-2283102

Dr. Pamela Christudoss  
CMCEQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not subcontract any components  
\*\*\*\*\*End of Report\*\*\*\*\*