



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12471		MEDICAL DEVICES	Page No:1 of 1	
Issue Date	30-09-2023			
SRF No & Date	SRF/23/00544-0001 & 29-09-2023			
Receipt Date	29-09-2023			
Calibration Date	29-09-2023			
Calibration Due	28-09-2024			
Customer Name & Address				
GOVERNMENT UPGRADE PRIMARY HEALTH CENTRE, KOMARALINGAM-642204.				
Details of Device Under Calibration (DUC)				
Description	: SEMI AUTO ANALYZER	Make & Model	: NA & PRIETEST TOUCH	
Range	: MULTI	Sr. No	: ATCD3100321RBK	
Resolution	: MULTI	Identification No	: NA	
DUC Condition	: SATISFACTORY	Location	: LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024


ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	80	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000μAfor B,BF,CF	195	7.4
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500μAfor B,BF,CF	219	7.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)
E.ESWAR

Authorised Signatory

 Technical Manager
C.SHANMUGARAJ
 Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.