

				CALI	BRATI	ON	CERT	<i>IFIC</i>	<u>A7</u>	E		
CERTIFICATE NO: SBS/CL/23/12472						MEDICAL DEVICES					Page	No:1 of 1
Issue Date						30-09-2023						
SRF No & Date						SRF/23/00544-0002 & 29-09-2023						
						29-09-2023						
						29-09-2023						
Calibration Due					28-09	-2024						
Custome	r Name	& Add	ress									
GOVERN	MENT U	PGRA	DE PRI	MARY HEALT	TH CENTRE	,						
KOMARA												
				I	Details of De	vice U	nder Calibrat	ion (DU	C)			
Description ELECTRICAL SAFE (MICROSCOPE)			Υ	Make	& Model			LABOMED & VISION	2008			
Range : MULTI				Sr. No)	:		221245973				
Resolution : MULTI				Identi	ication No			NA				
DUC Condition SATISFACTORY				Locat				LABORATORY				
BOO COM				Environr	nental Condi	tions	& Calibration	Proced	ure D	etails		
Environmental Details Temperature:			25.6°C Relative Hun		idity 52% RH							
Calibration Procedure No SBS/CP/MD/2			9 Calibration d		one at ONSITE							
- Cambration					Refere	ence S	andards Det	ails				
S.No Des	No Description				Make/ SI No:		Certificate No				Validity	
				Rigel Medical & 44L-1059		M-230809-16-4				10-08-2024		

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)		
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)		
	>20ΜΩ	93	13.92		
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)		
_	<5000µAfor B.BF,CF	154	7.7		
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)		
	<500µAfor B,BF,CF	209	7.3		

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

(Calibration Engineer) E.ESWAR

michiational otaliaa.	
Authorised	Signatory
Technical Manager	Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS