

Certificate No: PHC/MP/033

CERTIFICATE OF CALIBRATION

Customer Details	Details of DUC	
PRIMARY HEALTH CENTRE, MUNCHIRAI, VILAVANCODE, KANNIYAKUMARI, TAMILNADU- 629171	Description	MICROPIPETTE
	Manufactured by	NIL
	Model	100-1000µl
	Serial No.	RW03005
	Asset ID	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	27°C	Date of Receipt	26/09/2023
		Date of Calibration	28/09/2023
Humidity	56%	Calibration Due Date	27/09/2024
		Date of Issue	29/09/2023
		Cond. On Receipt	Satisfactory

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electronic Semi Micro Balance	AND, GR-202	14251938	27-DEC-2023	Equations

Performance Analysis

SI.No	Device Reading(µl)	Mean Value(µl)	Deviation
1.	1000	1000.1	0.1
2.	500	500.2	0.2
3.	100	99.9	0.1