



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, T.U.Mangai

[Details About Robust Analysis](#)

Lab No 7466

[Detail About Monthly Summary](#)

Month August

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	3.46	9.31	163.5 mg/dL	-1.14	2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13	1.8 mg/dL	0.32	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	110.2 mg/dL	-0.76	0.81

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out