



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. Block PHC, Palayampudur

Lab No

7351

Constituent Group Chemistry III

Date of Result Entered :

19/09/2023

PT item Lyophilized human serum based

Date of Report Published :

05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	327 mg/dL	1.08	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	112 mg/dL	-0.38	0.20
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ± 1.01 to ± 2.00		Good.								
Within ± 2.01 to ± 2.99		Accept with caution. Warning Signal.								
Beyond ± 3.0		Unacceptable performance. Action Signal.								

LAB ADDRESS :

Govt. Block PHC, Palayampudur
 THE MEDICAL OFFICER, PALAYAMPUR PHC, NALLAMPALLI, DHARMAPURI DIST.
 Palayampudur
 TAMILNADU636807

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****