

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name

Govt. PHC, Sevanthipatti

Lab No

12429

Constituent Group

Chemistry III

Date of Result Entered:

20/07/2023

PT item

Lyophilized human serum based

Date of Report Published:

07/08/2023

l.No	Analyte	Method / Principle	Analyzer Name	No of	ĐΥ	Partic	ipants	Your	SDI	9 (
		Name	Analyze: realise	Participants	O V	CV	SD	Value	JUI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	104 mg/dL	-0.88	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.2 mg/dL	1.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS: Govt. PHC, Sevanthipatti **CLINICAL LABORATORY** Sevanthipatti TAMILNADU621215

> **Coordinator Contact Details:** Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss

Dr. Pamela Christudoss **CMC EQAS Coordinator** Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report *****

> Mohudheka MEDICAL OFFICER Govt. Primary Health Centre, Seventhinalty - 621 215.