



V LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Komal [Details About Robust Analysis](#)
Lab No: 12400 [Detail About Monthly Summary](#)
Month: February [Detail about SDI](#)
Year: 2023
Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 21/02/2023
Date of Report Published: 08/03/2023

No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonk (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	140 mg/dL	-1.48	0.47
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robonk (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.4 mg/dL	0.00	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk (Semi automatic / Fully Automatic)	1768	112.79	5.23	5.90	125 mg/dL	2.07	0.28

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Monthly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Keral
Lab No: 12400
Month: March
Year: 2023
Constituent Group: Chemistry II

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023
Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants	
						CV	SD
1	GLUCOSE II	GOD-POD II	Robosnk (Semi automatic / Fully Automatic)	1878	91.48	7.48	6.82
2	CREATININE II	Jaffes Kinetic - Alkaline Phtrite II	Robosnk (Semi automatic / Fully Automatic)	1831	1.33	9.08	0.12
3	CHOLESTROL II	CHOD-PAP II	Robosnk (Semi automatic / Fully Automatic)	1701	110.93	5.23	5.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Lab Address	Lab Method	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Keral	Details About Robust Analysis
Lab No	12400	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robosnk (Semi automatic / Fully Automatic)	1832	115.43	8.04	7.21	130 mg/dL	1.46	0.34
2	CREATININE II	Jaffes Kinetic - Alkaline Picrate II	Robosnk (Semi automatic / Fully Automatic)	1620	2.16	8.09	8.18	2.3 mg/dL	1.54	0.01
3	CHOLESTROL II	CHOD-PAP II	Robosnk (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	130 mg/dL	1.45	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Labdy Address	Labdy Method	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Keral
Lab No: 12400
Month: May
Year: 2023
Constituent Group: Chemistry II

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/05/2023
Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robosik (Semi automatic / Fully Automatic)	1824	1.90.88	5.19	5.39	208 mg/dL	2.89	0.44
2	CREATININE II	Jaffes Kinetic - Alkaline Phthale II	Robosik (Semi automatic / Fully Automatic)	1824	1.76	7.48	8.13	1.0 mg/dL	-1.21	0.81
3	CHOLESTROL II	CHOD-PAP II	Robosik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.88	124 mg/dL	1.01	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Lab Address	Lab Method	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Keral
Lab No: 12400
Month: June
Year: 2023
Constituent Group: Chemistry II

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023
Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	EV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robosik (Semi automatic / Fully Automatic)	1823	295.63	3.17	14.82	200 mg/dL	0.23	0.89
2	CREATININE II	Jaffes Kinetic - Alkaline Phthale II	Robosik (Semi automatic / Fully Automatic)	1846	2.37	6.64	0.17	2.5 mg/dL	-0.41	0.61
3	CHOLESTROL II	CHOD-PAP II	Robosik (Semi automatic / Fully Automatic)	1804	113.73	4.48	5.07	110 mg/dL	-0.74	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Lab Address	Lab Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Komal	Details About Robust Analysis
Lab No	12400	Detail About Monthly Summary
Month	January	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/01/2023

Date of Report Published : 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Rebonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	116 mg/dL	-0.51
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Rebonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	1.7 mg/dL	-2.30
3	CHOLESTROL III	CHOD-PAP III	Rebonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	112 mg/dL	-0.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out