

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Mathigiri Lab No 12317

Constituent Group Date of Result Entered: 20/09/2023

PT item Lyophilized human serum based Date of Report Published: 05/10/2023

	SI.No	Amaluta	Method / Principle	Analyman Nama	No of	DV	Participants		Your Value	SDI	U	
31.10		o Analyte	Name	Analyzer Name	Participants	υV	CV	SD				
	1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	310 mg/dL	-0.06	0.69	
	2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00	
	3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	118 mg/dL	1.01	0.20	

SDI Range	Interpretation		
Within -1.00 to +1.00	Excellent.		
Within ±1.01 to ±2.00	Good.		
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.		
Beyond ±3.0	Unacceptable performance. Action Signal.		

LAB ADDRESS:

Govt. PHC, Mathigiri THE MEDICAL OFFICER, Hosur, Krishnagiri Mathigiri TAMILNADU635110

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******