



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13917

Page No: 1 of 1

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
NAKKAMBADI-612203, MAYILADUTHURAI DISTRICT.

SRF No.	SRF/23/00838-0003
SRF Date	09-10-2023
Date of Receipt	07-10-2023
Date of Calibration	09-10-2023
Due Date for Calibration	08-10-2024
Issue Date	10-10-2023

**Details of Unit Under Calibration**

Description	Micro Pipette	Make	THERMO FISHER
Range	10-100µl	Model	NA
Resolution	0.2µl	Material	PVC
Serial Number	RWD9941	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Calibration Method Used

**Standard Used**

Sl. No.	Description	ID.No. / SL No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10.0	9.47	9.46	9.48	9.47	9.46	9.45	-0.55	0.02	0.47
		9.42	9.43	9.43	9.44	9.45				
2	50.0	49.57	49.56	49.58	49.56	49.56	49.56	-0.44	0.01	0.47
		49.54	49.56	49.57	49.56	49.54				
3	100.0	98.92	98.94	98.93	98.94	98.93	98.91	-1.09	0.02	0.47
		98.91	98.92	98.89	98.88	98.88				

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

  
 (Calibration Engineer)  
 M.RAGUL



Authorised by:

  
 Quality Manager / Chief Executive  
 C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.