

CHRISTIAN MEDICAL COLLEGE  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - June 2023

Lab Name  
Constituent Group  
PT item

Govt. PHC, Indur  
Chemistry III  
Lyophilized human serum based

Lab No  
Date of Result Entered :  
Date of Report Published :

12113  
19/06/2023  
04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	268 mg/dL	-1.26	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	3.01 mg/dL	2.57	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	113.8 mg/dL	0.01	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Indur  
THE MEDICAL OFFICER, INDUR PHC, NALLAMPALLI, DHARMAPURI DIST.  
Indur  
TAMILNADU636803.

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

[Signature Box]  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*P. Christudoss*  
**MEDICAL OFFICER**  
GPHC  
INDUR (PO), NALLAMPALLI (TK)  
DHARMAPURI (DT) - 636 803.