

## CHRISTIAN MEDICAL COLLEGE

## DEPARTMENT OF CLINICAL BIOCHEMISTRY

## CMC EXTERNAL QUALITY ASSURANCE SCHEME

## **MONTHLY SUMMARY REPORT - AUGUST 2023**



Govt. PHC, Thenurpatty

Lab No

10559

**Constituent Group** 

Chemistry III

Date of Result Entered :

19/08/2023

PT item

Lyophilized human serum based

Date of Report Published:

01/09/2023

| 1 | GLUCOSE III    | GOD-POD III                       | Robonik ( Semi automatic / Fully<br>Automatic ) | 1841 | 175.58 | 5.13 | 9.00 | 182.2<br>mg/dL  | 0.74  | 0.42 |
|---|----------------|-----------------------------------|---|------|--------|------|------|-----------------|-------|------|
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate | Robonik ( Semi automatic / Fully<br>Automatic ) | 1852 | 1.76   | 6.77 | 0.12 | 2 mg/dL         | 2.02  | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III                      | Robonik ( Semi automatic / Fully<br>Automatic ) | 1764 | 113.80 | 4.07 | 4.64 | 110,7<br>mg/dt. | -0.67 | 0,22 |

| SDI Range             | Interpretation                           |  |  |  |
|-----------------------|--|--|--|--|
| Within -1.00 to +1.00 | Excellent.                               |  |  |  |
| Within ±1.01 to ±2.00 | Good.                                    |  |  |  |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |  |  |  |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |  |  |  |

LAB ADDRESS :

COVI. PHC, Thenurpatty
CLINICAL LABORATORY, NAMAKKAL(DIST)
Thenurpatty
TAMILNADU637411

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Punela Christados

TOTALON CONTRACTOR SECURITY