



Vanavil Calibrations

Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, SURULODU-629 164		Certificate No:	VCPL/CC/23-24/26-002
		Date of Calibration:	29-09-2023
		Date of issue:	30-09-2023
		Date of Receipt:	29-09-2023
		Next cal Date:	28-09-2024
Details of Device Under Calibration			
Description:	MICROPIPETTE	Make:	THERMOSCIENTIFIC
Range:	(10 to 100) μ l	Model/Type:	FINN PIPETTEF3
Resolution:	0.2 μ l	Condition of DUC	GOOD
Serial Number	RW02186	Location:	LABORATORY
ID number	-	Cal. At:	LAB
Environmental Conditions		Calibration Method Used	
Temperature	23.2 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	49 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

RESULT OF CALIBRATION


S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	10.0	9.88	0.12
2	100.0	99.87	0.13
3	1000.0	999.83	0.17

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

R.GOPALAKRISHNAN
(CALIBRATION ENGINEER)



Authorised by:

S. LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Mahara] Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

vanavilcalibrations@gmail.com www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com



Vanavil Calibrations

Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, SURULODU-629 164		Certificate No: VCPL/CC/23-24/26-003
		Date of Calibration: 29-09-2023
		Date of issue: 30-09-2023
		Date of Receipt: 29-09-2023
		Next cal Date: 28-09-2024
Details of Device Under Calibration		
Description:	MICROPIPETTE	Make: THERMOSCIENTIFIC
Range:	(100 to 1000) μ l	Model/Type: FINN PIPETTE F3
Resolution:	0.2 μ l	Condition of DUC: GOOD
Serial Number:	RW03045	Location: LABORATORY
ID number:	-	Cal. At: LAB
Environmental Conditions		Calibration Method Used
Temperature:	23.2 °C	Calibration Method Used: ISO 8655:2002 PART-6
Humidity:	49 % Rh	Cal procedure no: -

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024


RESULT OF CALIBRATION


S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100.0	99.87	0.13
2	500.0	499.87	0.13
3	1000.0	999.74	0.26

Remarks:

- 1.DUC is Defined as Device Under Calibration.
2. Calibration Frequency decided by the customer.
3. The above instrument is working in good condition.

Calibrated By:


R.GOPALAKRISHNAN
(CALIBRATION ENGINEER)



Authorised by:


S. LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

52, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com