

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS :		CERTIFICATE NO		TICS/2023/MM/SG/KAM/05	
GOVT.PRIMARY HEALTH CENTRE		EQUIPMENT NAME		MICROPIPETTE	
KALAIYARMANGALAM,		MANUFACTURE		MICROPET	
SIVAGANGAL.		MODEL		100-1000µL	
CALIBRATION DATE	03/10/2023	SERIAL NO	KAM PH 05		
CALIBRATION DUE DATE	02/10/2024	INSTALLED	LAB		
		DEPARTMENT			
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.1	± 1	PASS
		1000	999.3	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :		Authorized Sign With Seal			
					