

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Molesamangalam

Lab No 12954

Month March

Year 2023

Constituent Chemistry III  
Group

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)


Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

| Sl.No                 | Analyte        | Method / Principle Name               | Analyser Name                               | No of Participants                       | DV     | Participants CV | SD   | Your Value | SDI  | U    |
|-----------------------|----------------|---------------------------------------|---|--|--------|-----------------|------|------------|------|------|
| 1                     | GLUCOSE III    | GOD-POD III                           | Any Analyser (Automation / Semi Automation) | 74                                       | 92.52  | 7.71            | 7.13 | 93 mg/dL   | 0.07 | 1.66 |
| 2                     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Any Analyser (Automation / Semi Automation) | 70                                       | 1.32   | 7.52            | 0.10 | 1.36 mg/dL | 0.40 | 0.02 |
| 3                     | CHOLESTROL III | CHOD-PAP III                          | Any Analyser (Automation / Semi Automation) | 148                                      | 111.55 | 4.29            | 4.78 | 116 mg/dL  | 0.93 | 0.79 |
| SDI Range             |                |                                       |   | Interpretation                           |        |                 |      |            |      |      |
| Within -1.00 to +1.00 |                |                                       |   | Excellent.                               |        |                 |      |            |      |      |
| Within ±1.01 to ±2.00 |                |                                       |   | Good.                                    |        |                 |      |            |      |      |
| Within ±2.01 to ±2.99 |                |                                       |   | Accept with caution. Warning Signal.     |        |                 |      |            |      |      |
| Beyond ±3.0           |                |                                       |   | Unacceptable performance. Action Signal. |        |                 |      |            |      |      |

Links

|                                |                                 |                                    |  |
|--------------------------------|---------------------------------|------------------------------------|--|
| <a href="#">Modify Address</a> | <a href="#">Modify Methods</a>  | <a href="#">Submit Lab Results</a> | <a href="#">View Lab Monthly Summary</a> |
| <a href="#">Yearly Summary</a> | <a href="#">Change Password</a> | <a href="#">Home</a>               | <a href="#">Sign-out</a>                 |

  
 மூலச்சாரி சௌந்திராணி  
 மூலச்சாரி சௌந்திராணி  
 மூலச்சாரி சௌந்திராணி  
 மூலச்சாரி சௌந்திராணி - 604504