

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Melsesamangalam  
 Lab No 12954  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52	191 mg/dL	0.93	2.0
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12	1.85 mg/dL	0.81	0.4
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	127 mg/dL	2.09	0.9

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

மருத்துவ அலுவலர்  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 மேல்சீமங்கலம்,  
 செய்யார் தாலுக்கா - 60450

links

me.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2\_new.aspx?resDetId=265&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110