

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Melsesamangalam
 Lab No 12954
 Month July
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95	109 mg/dL	-0.12	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27	5.85 mg/dL	-0.15	0.07
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88	99 mg/dL	-2.82	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

kei
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்,
 மேல்சீமங்கலம்,
 செய்யார் தாலுக்கா - 604504

Links

s://home.cmcvellore.ac.in/dinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=267&noSamples=1&month=7&year=2023&groupId=9&yearGroup=110