

MONTH EQAS PARAMETERS VALUE SDF RANGE Mo SILVA

SEPTEMBER  
23

EQAS PARAMETERS	VALUE	SDF RANGE
Glucose	286	-1.66
Cholesterol	116	0.55
Creatinine	1.5	2.00

CAPA

- \* Glucose and cholesterol - SDF Range within limit
- \* Creatinine - SDF Range - Accept within warning signal Caution
- \* Called BME - maintain SAA cleaning and washing
- \* Re calibrated Creatinine test, Rerun PqAS and EqAS

PqAS - Creatinine value - 1.35  
EqAS - Creatinine value - 1.2

*[Signature]*  
Officer  
Government Primary Health Centre  
Guruvarejappet - 685101  
Kannur District