

**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. Block PHC, Thookanickanpalayam  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based  
 Date of Result Entered: 15/05/2023  
 Date of Report Published: 05/06/2023

Lab No: 7377

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	167 mg/dL	-1.48	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik ( Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.71 mg/dL	-0.38	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	104.9 mg/dL	-1.66	0.27

SDI Range		Interpretation
Within -1.00 to +1.00	Excellent.	
Within ±1.01 to ±2.00	Good.	
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.	
Beyond ±3.0	Unacceptable performance. Action Signal.	

LAB ADDRESS :  
 Govt. Block PHC, Thookanickanpalayam  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT.  
 Thookanickanpalayam  
 TAMILNADU638506

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

*Pamela Christudoss*  
 சர்க்கரியில் சர்க்கரியில் சர்க்கரியில்  
 சர்க்கரியில் சர்க்கரியில் சர்க்கரியில்

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential