




CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. Block PHC, Thookanaikanpalayam

Lab No: 7377

Constituent Group: Chemistry III

Date of Result Entered: 15/04/2023

PT Item: Lyophilized human serum based

Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	115 mg/dL	-0.62	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2 mg/dL	-0.91	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	104 mg/dL	-1.31	0.28
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Handwritten signature

மருத்துவ அலுவலர்,
சுகாதார அலுவலர்
சுகாதார அலுவலர்

Handwritten signature: Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Handwritten signature

AB ADDRESS :
Govt. Block PHC, Thookanaikanpalayam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT.
hookanaikanpalayam
AMILNADUG638506

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential