



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. Block PHC, Thookanaickanpalayam

Lab No: 7377

Constituent Group: Chemistry III

Date of Result Entered: 12/09/2023

PT Item: Lyophilized human serum based

Date of Report Published: 05/10/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	331.1 mg/dL	1.35	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.42 mg/dL	1.24	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	120 mg/dL	1.48	0.20
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. Block PHC, Thookanaickanpalayam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT.
Thookanaickanpalayam
TAMILNADU638506

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

மேல்க்கண்ட தரவுகளைப் பற்றி
மேல்க்கண்ட தரவுகளைப் பற்றி
மேல்க்கண்ட தரவுகளைப் பற்றி

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential