

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Melakodumalur		<a href="#">Details About Robust Analysis</a>							
Lab No	12564		<a href="#">Detail About Monthly Summary</a>							
Month	June		<a href="#">Detail about SDI</a>							
Year	2023									
Constituent Group	Chemistry III									
<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 19/06/2023							
			Date of Report Published : 04/07/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Any Analyser (Automation / Semi Automation )	65	287.55	4.50	12.95	287.68 mg/dL	0.01	3.21
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	67	2.58	6.39	0.16	2.48 mg/dL	-0.61	0.04
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	149	114.61	4.05	4.64	122 mg/dL	1.59	0.76
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>