

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Melakodumalur	<a href="#">Details About Robust Analysis</a>									
Lab No	12564	<a href="#">Detail About Monthly Summary</a>									
Month	May	<a href="#">Detail about SDI</a>									
Year	2023										
Constituent Group	Chemistry III										
<a href="#">Click on the analyte to view Graphical Data</a>										<a href="#">All Analyser Result</a>	<a href="#">Print</a>
										Date of Result Entered : 19/05/2023	
										Date of Report Published : 05/06/2023	
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Any Analyser (Automation / Semi Automation )	69	181.20	5.81	10.52	174.39 mg/dL	-0.65	2.53	
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	68	1.75	7.05	0.12	1.72 mg/dL	-0.24	0.03	
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	154	114.84	5.08	5.83	118.81 mg/dL	0.68	0.94	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>