

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Melakodumalur	<a href="#">Details About Robust Analysis</a>								
Lab No	12564	<a href="#">Detail About Monthly Summary</a>								
Month	March	<a href="#">Detail about SDI</a>								
Year	2023									
Constituent Group	Chemistry III									
<a href="#">Click on the analyte to view Graphical Data</a>							<a href="#">All Analyser Result</a>		<a href="#">Print</a>	
Date of Result Entered : 20/03/2023										
Date of Report Published : 04/04/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Any Analyser (Automation / Semi Automation )	74	92.52	7.71	7.13	89.95 mg/dL	-0.36	1.66
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	70	1.32	7.52	0.10	1.29 mg/dL	-0.30	0.02
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	148	111.55	4.29	4.78	113.32 mg/dL	0.37	0.79
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>