



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/15805		MEDICAL DEVICES		Page No:1 of 1
Issue Date	26-10-2023			
SRF No & Date	SRF/23/00954-0001 & 25-10-2023			
Receipt Date	25-10-2023			
Calibration Date	25-10-2023			
Calibration Due	24-10-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, PUDUR-626905, THOOTHUKUDI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIESTEST TOUCH	
Range	MULTI	Sr. No	ATCD1011010RBK	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	BIO CHEMISTRY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature 25.6 °C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 99	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B, BF, CF	Measured values in µA 154	Uncertainty in % (±) 7.7
3	Enclosure Leakage <500µA for B, BF, CF	Measured values in µA 211	Uncertainty in % (±) 7.3

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

E. Q.
(Calibration Engineer)
E.ESWAR



Authorised Signatory

C. Shw.
Technical Manager
C SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkartuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.