



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/15809

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,  
PUDUR-628905, THOOTHUKUDI DISTRICT.

SRF No.	SPT/23/001/2A-0005
SRF Date	25-10-2023
Date of Receipt	23-10-2023
Date of Calibration	25-10-2023
Due Date for Calibration	24-10-2024
Issue Date	26-10-2023

**Details of Unit Under Calibration**

Description	MICRO PIPETTE	Make	DRAGON LAB
Range	100-1000µl	Model	NA
Resolution	10µl	Material	PVC
Serial Number	YE225AX025836SCE	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	BLOOD BANK

**Environmental Condition**

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Calibration Method Used**

**Standard Used**


SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP/L22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


Z Factor: 1.00319


Result of Calibration in µl											
SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.77	99.78	99.76	99.75	99.74	99.77	-0.23	0.07	0.47	
		99.75	99.76	99.72	99.71	99.95					
2	500	499.95	499.96	499.95	499.96	499.97	499.96	-0.04	0.01	0.47	
		499.96	499.97	499.98	499.98	499.96					
3	1000	999.88	999.87	999.88	999.86	999.87	999.88	-0.12	0.01	0.47	
		999.89	999.87	999.88	999.89	999.87					

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By:   
(Calibration Engineer)  
M.RAGUL



Authorised by:   
(Quality Manager/Chief Executive)  
C.SIVABALAN

### SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.