

**CHRISTIAN MEDICAL COLLEGE**DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023

PC-1024

Lab Name Govt. PHC, Kalikulam

Lab No

12927

Constituent Group Chemistry III

Date of Result Entered :

19/08/2023

PT Item Lyophilized human serum based

Date of Report Published :

01/09/2023

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	165 mg/dL	-1.18	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.4 mg/dL	-3.03	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112.08 mg/dL	-0.37	0.22
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Kalikulam
Kazhikulam
Kalikulam
TAMILNADU

Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102*Pamela Christodoss*Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore