

**CHRISTIAN MEDICAL COLLEGE**

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. PHC, Kalikulam

12927

Constituent Group Chemistry III

Lab No

20/03/2023

Date of Result Entered :

PT item Lyophilized human serum based

Date of Report Published :

04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	92 mg/dL	0.08	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	118 mg/dL	1.22	0.28
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent						
Within ± 1.01 to ± 2.00				Good.						
Within ± 2.01 to ± 2.99				Accept with caution. Warning Signal.						
Beyond ± 3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Kalikulam

Kazhikulam

Kalikulam

TAMILNADU

Coordinator Contact Details:Email: clingc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore