



Vanavil Calibrations

Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/s Government Primary Health Centre Pethappampatty-6422205	Report No	VCPL/2023/002-6
	Date of Receipt	23/Mar/2023
	Date of Issue	24/Mar/2023
	Calibration Date	23/Mar/2023
	Next Due Date	22/Mar/2024
	Calibration at	On Site

Details of Device Under Calibration:

Description:	MICRO PIPETTE	Make :	MICROLUX
Range:	100 to 1000 µl	Model/Type:	-
Resolution:	10 µl	Accuracy :	As per Manual
Serial Number:	-	Location:	LAB
ID number:	GPH/PIP/02	Condition of DUC:	GOOD

Temperature	22.8 °C	National/International standard:	-
Humidity	52 % Rh	Cal procedure no:	-

Master Details


SL no	Description	Serial No	Certificate no	Valid Till
1	ELECTRONIC BALANCE	16031172	E150323-1	15-Mar-24

CALIBRATION RESULT

S.NO	DUC Readings (µl)	Standard Rreadings Mean Value (mg)	ERROR (mg)
1	100	99.90	-0.10
2	500	499.80	-0.20
3	1000	999.76	-0.24


- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


T.SUDHAHAR
(CALIBRATION ENGINEER)



Authorised by:


S. LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

