

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab name Govt. PHC,
 Katchaidalavaipuram
 Constituent Chemistry III
 Group
 PT item Lyophilized human serum
 based

Lab No 10635
 Date of Result Entered : 20/08/2023
 Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	176.7 mg/dL	0.12	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.75 mg/dL	-0.08	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	118.8 mg/dL	1.08	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Katchaidalavaipuram
 CLINICAL LABORATORY, TUTICORIN District .,

MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KATCHERITHALAVAIPURAM
 THOOTHUKUDI DIST